K123760

ASSURGUARD SDN. BHD. (Company No.888413-H)

FDA 510(k), Premarket Notification : 510(k) Summary of Safety and Effectiveness Information

Date: 16 January 2013

JAN 28 2013

1.0 Submitter:

Name:	ASSURGUARD SDN. BHD.
Address:	82F, Jalan Pulasan, 41000 Klang, Selangor Darul Ehsan, Malaysia.
Country:	Malaysia
Phone No.:	+603 3297 1020 Fax No.: +603-3291 3594
Registration No.:	Pending (First Device)

2.0 Contact Person:

Contact:	Mr. Lim Hui Guan
E-mail:	assurguard@gmail.com
Telephone No.:	+603 3297 1020
Fax No.:	+603 3291 3594

3.0 Name of Device:

Trade Name:	Powder Free Nitrile Patient Examination Gloves, Blue Colored and White (Non-Colored). Non-Sterile.
Common Name:	Patient Examination Glove
Classification Name:	Patient Examination Glove.

4.0 Identification of The Legally Marketed Device:

The Powder Free Nitrile Patient Examination Gloves, Blue Colored and White (Non-Colored), Non-Sterile, Class I Patient Examination gloves, 80LZA, meets all of the requirements of ASTM D6319 Standard Specification for Nitrile Examination Gloves for Medical Application.

Predicate Device: K121464, Blue Powder Free Nitrile Patient Exam Glove

5.0 Description of Device:

Powder Free Nitrile Patient Examination Gloves, Blue Colored and White (Non-Colored), Non-Sterile meet all the current specification for ASTM D6319.

6.0 Intended Use of the Device:

A patient examination glove is a disposable device intended for medical purposes that is worn on the examiner's hand to prevent contamination between patient and examiner.

7.0 Summary of The Technological Characteristics of The Device:

Powder Free Nitrile Patient Examination Gloves, Blue Colored and White (Non-Colored), Non-Sterile, possesses the following technological characteristic (as compared to ASTM or equivalent standards):

Characteristic	Standards	Device Performance
Dimensions	ASTM D 6319-10	Meets
Physical Properties	ASTM D 6319-10	Meets
Freedom from pin-holes	ASTM D 5151-11 ASTM D 6319-10	Meets Meets
Powder Free Residue	ASTM D 6124-11 ASTM D 6319-10	Meets Meets
Biocompatibility	Dermal Sensitization (as per ISO 10993-10:2010)	Not a contact skin sensitizer
	Primary Skin Irritation Test (as per 16 CFR Part 1500)	Not a primary skin irritant

8.0 Substantial Equivalent Based on Assessment of Non-Clinical Performance Data

The performance test data that support a determination of substantial equivalence are described above.

9.0 Substantial Equivalent Based on Assessment of Clinical Performance Data

Clinical data are not needed for market cleared examination gloves.

10.0 Conclusion

It can be concluded that the Powder Free Nitrile Patient Examination Gloves, Blue Colored and White (Non-Colored), Non-Sterile is safe and effective for use and will perform according to the glove performance standards referenced in Section 7.0 above.

11.0 Substantial Equivalence Discussion

There is no different between the proposed device and the predicate with respect to indications for use and technological characteristics.

The gloves are identical to predicate device with 510(k) K121464.

The substantial equivalence comparison is presented in Table below:-

Section 11.0 Substantial Equivalence Comparison

Characteristics	Predicate Device K121464, Blue Powder Free Nitrile Patient Exam Glove	Proposed Device Powder Free Nitrile Patient Examination Gloves, Blue Colored and White (Non-Colored). Non-Sterile.
Product Code	80 LZA	Same
FDA Device Class	Class I	Same
Intended Use	Intended for medical purposes that is worn on the examiner's hand to prevent contamination between patient and examiner.	Same
Indications for Use	A patient examination glove is a disposable device intended for medical purposes that is worn on the examiner's hand to prevent contamination between patient and examiner.	Same
Construction	Ambidextrous, Powder Free, **Blue,** per ASTM D6319 specification.	Ambidextrous, Powder Free, **Blue and White**, per ASTM D6319 specification.
Materials	Nitrile Butadiene Rubber	Same
Performance I. Sterility II. Freedom from holes III. Dimension	 Non-Sterile Meets ASTM D6319 Meets ASTM D6319	 Same Same Same

IV. Physical Properties V. Powder Free Residue	Meets ASTM D6319 Meets ASTM D6319	Same Same
Single Use	Yes	Same
Biocompatibility Test	Passes i. Primary Skin Irritation Test ii. Dermal Sensitization Test	Same
Packaging	Packed in Dispenser Boxes	Same

DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

January 28, 2013

Mr. Lim Hui Guan
Managing Director
Assurguard Sdn. Bhd.
82F, Jalan Pulasan
Selangor Darul Ehsan
Malaysia 41000

Re: K123760
Trade/Device Name: Powder Free Nitrile Patient Examination Gloves, Blue Colored and White (Non-Colored). Non-Sterile
Regulation Number: 21 CFR 880.6250
Regulation Name: Patient Examination Glove
Regulatory Class: I
Product Code: LZA
Dated: December 1, 2012
Received: December 7, 2012

Dear Mr. Guan:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm.

Sincerely yours,

Anthony D. Watson, B.S., M.S., M.B.A.
Director
Division of Anesthesiology, General Hospital,
Respiratory, Infection Control and
Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K123760

Device Name: Powder Free Nitrile Patient Examination Gloves, Blue Colored and White (Non-Colored). Non-Sterile

Indication For Use:

A patient examination glove is a disposable device intended for medical purposes that is worn on the examiner's hand to prevent contamination between patient and examiner.

Prescription Use _____ AND/OR Over-The-Counter Use X
(Part 21 CFR 801 Subpart D) (21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office Of Device Evaluation (ODE)

Elizabeth F. Claverie
2013.01.22 18:10:26 -05'00'

(Division Sign-Off)
Division of Anesthesiology, General Hospital Infection Control, Dental Devices

510(k) Number: K123760